FARMER FREEMAN

1415 Elliot Pl NW, Building B, Washington, DC 20007
202-697-7640 support@farmerfreeman.com
www.farmerfreeman.com

Pathogen Diagnostic Report

LCV

CUSTOMER NAME
Morgan Chen

COMPANY / ORGANIZATION
Farmer Freeman

EMAIL
support@farmerfreeman.com

REPORT DATE
October 15, 2026

RUN DATE
October 15, 2026

DATE OF LAB RECEIPT
October 14, 2026

Diagnostic Results

Sample ID	Strain	Nickname	Sample Type	LCV	IC
VT00001	Gorilla Glue #4	GG4-A01	Root + Leaf	NOT DETECTED	PASS 26.20
VT00002	Girl Scout Cookies	GSC-B02	Root + Leaf	DETECTED 24.80	N/A 25.10
VT00003	Zkittlez	ZK-C03	Root + Leaf	NOT DETECTED	PASS 24.90
VT00004	Runtz	RZ-D04	Root + Leaf	NOT DETECTED	PASS 25.70
VT00005	Sunset Sherbet	SS-E05	Root + Leaf	NOT DETECTED	PASS 24.30

- DETECTED — Pathogen present; IC not required for detection
- NOT DETECTED — No pathogen signal; IC valid
- UNDER ANALYSIS — IC failed; being retested, ready in 1–2 business days
- Lower Cq values typically correlate with a greater pathogen load in the sample.

Interpretation Summary

Pathogen Detected

One or more samples tested positive for: **LCV**. Immediate action is recommended. Isolate affected plants and equipment. Do not take cuttings from positive plants. Sanitize tools between plants. Contact a cultivation advisor to develop a remediation plan.

About the Pathogens Tested

Lettuce Chlorosis Virus (LCV)

Channel: ROX **Type:** Virus **Threshold:** Cq < 35

Lettuce Chlorosis Virus (LCV) is a crinivirus transmitted by whiteflies. It causes interveinal chlorosis, leaf curling, and reduced plant vigor. While less studied in cannabis than HLVd, LCV can reduce plant health and yield.

Methodology


This report utilizes quantitative reverse transcription PCR (RT-qPCR) performed on an Agilent AriaMX Real-Time PCR instrument. Nucleic acids are extracted from plant tissue using a validated extraction protocol. Each reaction includes an internal extraction and amplification control (IC/HEX channel) to confirm assay validity. A Cq threshold of 35 cycles is applied for most targets (40 cycles for Russet Mite): samples with Cq at or below threshold are reported as Detected. Samples with Cq between 35 and 40 cycles are also reported as Detected when the IC Cq is below 35, confirming assay validity. A positive target result is reported as Detected regardless of IC status. Wells with no target signal and a failed IC are reported as Under Analysis and will be retested by the laboratory.

Questions? Contact support@farmerfreeman.com or call 202-697-7640. Submit new samples at www.farmerfreeman.com.

VT00001

Gorilla Glue #4

GG4-A01



No pathogen detected
Pathogen detected
Not tested

PANEL
LCV

SAMPLE TYPE
Root + Leaf

DATE OF RECEIPT
October 14, 2026

RUN DATE
October 15, 2026


Results

LCV	**NOT DETECTED**	No Cq
Internal Control (IC)	**PASS**	Cq 26.20

VT00002

Girl Scout Cookies

GSC-B02



No pathogen detected
Pathogen detected
Not tested

PANEL
LCV

SAMPLE TYPE
Root + Leaf

DATE OF RECEIPT
October 14, 2026


RUN DATE
October 15, 2026

Results

LCV	**DETECTED**	Cq 24.80
Internal Control (IC)	**PASS**	Cq 25.10

VT00003

Zkittlez
ZK-C03



No pathogen detected
Pathogen detected
Not tested

PANEL
LCV

SAMPLE TYPE
Root + Leaf

DATE OF RECEIPT
October 14, 2026


RUN DATE
October 15, 2026

Results

LCV	**NOT DETECTED**	No Cq
Internal Control (IC)	**PASS**	Cq 24.90

VT00004

Runtz
RZ-D04



No pathogen detected
Pathogen detected
Not tested

PANEL
LCV

SAMPLE TYPE
Root + Leaf

DATE OF RECEIPT
October 14, 2026

RUN DATE
October 15, 2026


Results

LCV	**NOT DETECTED**	No Cq
Internal Control (IC)	**PASS**	Cq 25.70

VT00005

Sunset Sherbet

SS-E05



No pathogen detected
Pathogen detected
Not tested

PANEL
LCV

SAMPLE TYPE
Root + Leaf

DATE OF RECEIPT
October 14, 2026

RUN DATE
October 15, 2026

Results

LCV	**NOT DETECTED**	No Cq
Internal Control (IC)	**PASS**	Cq 24.30